

KaVo PiezoLED™ Ultrasonic Scaler Adjustment Recommendation

| Product Code | Therapeutic Indication | Power Setting | Spray Water Volume | Mode Adjustment* |
|--|------------------------|---------------|--------------------|------------------|
| PIEZO Scaler Tips Scaler tip 201 Scaler tip 202 Scaler tip 203 | | | | P3 P2 P1 (E) |
| | | | | P3 P2 P1 (E) |
| | | | | P3 P2 P1 (E) |
| PIEZO Paro Tips Paro tip 210 + 211 Paro tip 212 + 213 Paro tip 214 | | | | P3 P2 P1 (E) |
| | | | | P3 P2 P1 (E) |
| | | | | P3 P2 P1 (E) |

recommended
 possible
 not recommended

* Only applicable for ESTETICA treatment units and Primus 1058 Life. PiezoLED for Primus 1058 is bound to P3.

KaVo PiezoLED™ Ultrasonic Scaler

| PIEZO Prep Tips | Product Code | Therapeutic Indication | Power Setting | Spray Water Volume | Mode Adjustment* |
|-----------------|--------------------|------------------------|---------------|--------------------|------------------|
| | CEM tip 225 | | | | (P3) (P2) (P1) E |
| | Prep tip 226 | | | | P3 P2 P1 (E) |
| | Prep tip 227 + 228 | | | | P3 P2 P1 (E) |
| | Prep tip 229 | | | | P3 P2 P1 (E) |

| PIEZO Endo Tips | Product Code | Therapeutic Indication | Power Setting | Spray Water Volume | Mode Adjustment* |
|-----------------|-------------------------|--|---------------|--------------------|---|
| | Endo tip 220 | | | | P3 P2 P1 (E) |
| | Endo tip 221 | Tip not applicable for Primus 1058 | | | P3 P2 P1 E |
| | Endo tip 222 with files | Tip not applicable for Primus 1058 | | | P3 P2 P1 E |

| PIEZO Implant Tips | Product Code | Therapeutic Indication | Power Setting | Spray Water Volume | Mode Adjustment* |
|--------------------|-----------------|------------------------|---------------|--------------------|------------------|
| | Implant tip set | | | | P3 P2 P1 (E) |

recommended
 possible
 not recommended

* Only applicable for ESTETICA treatment units and Primus 1058 Life. PiezoLED for Primus 1058 is bound to P3.